

Supplementary Materials

A Low-Cost Micro-Volume Nephelometric System for Quantitative Immunoagglutination Assays

Qiqi Sun ^{1,2,*}, Wei Zheng ¹, Chao Lin ² and Dongxuan Shen ²

¹ Research Laboratory for Biomedical Optics and Molecular Imaging, Shenzhen Key Laboratory for Molecular Imaging, Shenzhen Institutes of Advanced Technology, Chinese Academy of Sciences, Shenzhen 518055, China; zhengwei@siat.ac.cn

² Edan Instruments, Inc., Shenzhen 518067, China; linchao@edan.com.cn (C.L.); dshen@edan.com.cn (D.S.)

* Correspondence: qsunaa@connect.ust.hk



Figure S1. Photograph of the nephelometric system.

Table S1. Results of experiments to evaluate the precision of system.

Experiment No.	Signal Change Rate (s ⁻¹)
1	0.000307782
2	0.000318123
3	0.00029997
4	0.000316522
5	0.000329339
Average	0.000314347
STD	1.11079E-05
C.V.	0.035336254